

Moisture Hammer Probe

Introduction


Congratulations on your purchase of the Extech Hammer Probe. This probe, when connected to an MO290 Moisture Meter, measures moisture in wood and other building materials. With proper care, this device will provide years of reliable service.

Safety Instructions

- This device should be used only for the purposes outlined in this guide. Other uses may cause damage to the probe and/or external devices, and can cause personal injury.
- Keep the probe packed when not in use.
- Do not allow children to handle the probe.

Description

1. Probe pins
2. Probe nuts
3. Hammer handle
4. Probe cable
5. RJ-45 connector



Operating Instructions

- Carefully remove the probe and two pins from the packaging box (extra pins are provided).
- Loosen and remove the nuts from the probe and insert the flat end of the pins into the holders. Slide the nuts over the pins and tighten.
- Connect the probe's RJ-45 connector to the socket at the bottom of the MO290 meter.
- Position the pins on the area under test.
- Drive the pins into the material under test using the hammer action force to assist.
- Operate the MO290 as described in the User Guide supplied with the meter.

Specifications

Cable Length	30" (0.8m) cable length
Hammer length	13" (33cm) without pins inserted
Pin length	1.6" (4cm)
Weight	4.2 lbs. (1.9kg)
Replacement pins	Part No.MO290-PINS-HP

Copyright © 2010 Extech Instruments Corporation (a FLIR company) - www.extech.com - V1.0 7/10

All rights reserved including the right of reproduction in whole or in part in any form.

airconcern

more info for Extech MO290-HP

Phone: 01235 838 555

Email: cs@airconcern.co.uk

Web: www.airconcern.co.uk

Air Concern Ltd, Building 173 Curie Avenue Harvell Didcot, Oxfordshire