

Carbon Dioxide 0.1%/a

Catalog No CH23501

Standard Measuring Range: 0.5 to 6 / 0.1 to 1.2 Vol %
Number of Strokes (n): 1 / 5
Time for Measurement: about 30 seconds / about 2.5 minutes
Standard Deviation: ± 5 to 10%
Color Change: white to pale violet to blue violet

Ambient Operation Conditions

Temperature: 0 to 30°C
Absolute Humidity: max 30 mg H₂O/L

Reaction Principle

CO₂ + Amine → blue violet reaction product

Cross Sensitivity

Hydrogen sulfide and sulfur dioxide in the TLV range do not interfere.



Manufactured by Draeger Safety
Distributed By Air Concern Ltd.

AIR CONCERN
Cleaner air quality life

www.AirConcern.co.uk • sales@AirConcern.co.uk

Dräger